working with phosphorus-33 safely

Radioactive half-life T_{1/2}

Principal emission

Monitoring for contamination

Biological monitoring

20 mSv annual limit on intake by inhalation

Shielding required

Special considerations

• Always follow the ten golden rules.

25.6 days

0.249 MeV beta (maximum)

Beta detector

Urine samples

 1.4×10^7 Bg (~ 0.38 mCi)

1-cm perspex/plexiglas. Although thinner shielding is adequate to reduce dose, it does not have good mechanical properties.

Half-life 25.6 days

	1.276	1.242	1.209	1.176	1.145	1.114	1.085	1.056	1.027	1.000
days	1 0.973	2 0.947	0.922	0.897	5 0.873	6 0.850	7 0.827	0.805	9 0.784	10 0.763
days	11	12	13	14	15	16	17	18	19	20
	0.742	0.723	0.703	0.685	0.666	0.648	0.631	0.614	0.598	0.582
days	21	22	23	24	25	26	27	28	29	30
	0.566	0.551	0.537	0.522	0.508	0.495	0.481	0.469	0.456	0.444
days	31	32	33	34	35	36	37	38	39	40
	0.432	0.421	0.409	0.398	0.388	0.377	0.367	0.357	0.348	0.339
days	41 0.330	42 0.321	43 0.312	0.304	45 0.296	46 0.288	47 0.280	48 0.273	49 0.265	50 0.258
days	51	52	53	54	55	56	57	58	59	60
	0.251	0.245	0.238	0.232	0.226	0.220	0.214	0.208	0.202	0.197
days	61	62	63	64	65	66	67	68	69	70
	0.192	0.187	0.182	0.177	0.172	0.168	0.163	0.159	0.154	0.150
days	71	72	73	74	75	76	77	78	79	80
	0.146	0.142	0.139	0.135	0.131	0.128	0.124	0.121	0.118	0.115
days	81	82	83	84	85	86	87	88	89	90
	0.112	0.109	0.106	0.103	0.100	0.097	0.095	0.092	0.090	0.087
days	91	92	93	94	95	96	97	98	99	100
	0.085	0.083	0.081	0.079	0.076	0.074	0.072	0.070	0.069	0.067

33 **D**

The data provided is general information that gives a basic understanding of radiation safety. You must however consult your local radiation safety expert to ensure that you comply with all national regulations and local rules. All numbers are taken from The Radionuclide and Radiation Protection Data Handbook 2002, Radiation Protection Dosimetry, Vol 98(1), Nuclear Technology Publishing, (2002).



10 golden rules

rule other considerations 1. Understand the nature of the Never work with unprotected cuts or breaks in the skin, particularly on the hands or hazard and get practical training. forearms. Never use any mouth-operated equipment in any area where unsealed radioactive material is used. Always store compounds under the conditions recommended. Label all containers clearly, indicating nuclide, compound, specific activity, total activity, date, and name of user. Containers should be properly sealed. 2. Plan ahead to minimize time Carry out a dummy run without radioactivity to check your procedures (the shorter spent handling radioactivity. the time, the smaller the dose). 3. Distance yourself appropriately Doubling the distance from the source quarters the radiation dose (The Inverse from sources of radiation. Square Law). 4. Use appropriate shielding for the 1-cm perspex/plexiglas will stop all beta particles but it is important to be aware of type of radiation. Bremsstrahlung from high-energy beta-emitters. Use suitable thickness of lead or lead acrylic shielding for X-ray and γ emitters. 5. Contain radioactive materials Always keep active and inactive work separated as far as possible, preferably by within defined work areas. maintaining rooms used solely for radioactive work. Always work over a spill tray within a ventilated enclosure. These rules may be relaxed for small (a few tens of kBq) quantities of ³H-, ³⁵S-, ³³P-, ¹⁴C-, and ¹²⁵I-labelled compounds in a non-volatile form in solution. 6. Wear appropriate protective Laboratory overalls, safety glasses, and surgical gloves must be worn at all times. clothing and dosimeters. However, beware of static charge on gloves when handling fine powders. Local rules will define which dosimeters should be worn (e.g. body film badge or thermo-luminescent extremity dosimeter for work with high energy beta-emitters). 7. Monitor the work area frequently In the event of a spill follow the prepared contingency plan: for contamination control. i. Verbally warn all people in the vicinity ii. Restrict unnecessary movement into and through the area iii.Report the spill to the Radiation Protection Supervisor/Adviser iv.Treat contaminated personnel first v. Follow clean-up protocol. 8. Follow the local rules and safe Do not eat, drink, smoke, or apply cosmetics in an area where unsealed radioactive ways of working. substances are handled. Use paper wipes and dispose of them appropriately. Never pipette radioactive solutions by mouth. Always work carefully and tidily. Minimize accumulation of waste Use the minimum quantity of radioactivity needed for the investigation. Disposal of all and dispose of it by appropriate radioactive waste is subject to statutory control. Be aware of the requirements and use only authorized routes of disposal. routes.



GE, imagination at work, and GE Monogram are trademarks of General Electric Company. All goods and services are sold subject to the terms and conditions of sole of the company within GE Healthcare which supplies them. GE Healthcare reserves the right, subject to any regulatory and contractual approval, if required, to make changes in specific actions and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your local GE Healthcare representative for the most current information. © 2006 General Electric Company—All rights reserved.